

Instructions for Use  
**Slimplicity® Solo Anterior Cervical Plate System**

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician

**DISCLAIMER OF WARRANTY AND LIMITATION OF REMEDY**

There is no express or implied warranty, including any implied warranty of merchantability or fitness for a particular purpose, on **Precision Spine®** product(s) described in this publication. Under no circumstances shall **Precision Spine** be liable for any direct, incidental or consequential damages other than as expressly provided by specific law. No person has the authority to bind **Precision Spine** to any representation or warranty except as specifically set forth herein.

Descriptions or specifications in **Precision Spine** printed matter, including this publication, are meant solely to generally describe the product at the time of manufacture and do not constitute any express warranties.

**DEVICE DESCRIPTION:**

**The Slimplicity Solo Anterior Cervical Plate System** consists of an assortment of plates and screws. Each Slimplicity Solo plate attaches to the anterior portion of the vertebral body of the cervical spine (levels C2-C7). Slimplicity Solo implants are composed of titanium alloy, as specified in ASTM F136. The components will be provided non-sterile.

**INDICATIONS:**

**The Slimplicity Solo Anterior Cervical Plate System** is intended for anterior cervical fixation to the cervical spine C2-C7 for the following indications: degenerative disc disease (as defined by neck pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), trauma (including fractures), tumors, deformity (defined as kyphosis, lordosis, or scoliosis), pseudarthrosis, failed previous fusion, spondylolisthesis, and spinal stenosis.

**PRECAUTIONS:**

Based on the dynamic testing results, the physician should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc., which may impact on the performance of the intervertebral body fixation device. The implantation of the intervertebral body fixation device should be performed only by experienced spinal surgeons with specific training in the use of this device because this is a technically demanding procedure presenting a risk of serious injury to the patient.

**CONTRAINDICATIONS:**

**The Slimplicity Solo Anterior Cervical Plate System** contraindications include, but are not limited to:

1. Patients with infection in or adjacent to the spine or spinal structures
2. Inadequate tissue coverage over operative site
3. Patients with morbid obesity
4. Pregnancy
5. Bone absorption, rapid joint disease, osteomalacia, osteopenia, and/or osteoporosis
6. Any spinal surgery case not needing a fusion
7. Any reuse, or multiple use
8. Fever or leukocytosis
9. Any patient unwilling or resistant to following postoperative instructions
10. Mental Illness
11. Cardiovascular complications
12. Allergic or other reaction to the metallic components and/or implants

**POTENTIAL ADVERSE AFFECTS:**

The following potential adverse effects associated with the procedure have been shown to occur with the use of similar spinal systems. All patients considered candidates for fusion should be informed concerning the pathogenesis of their spinal abnormality, the rationale for fusion with instrumentation, and the potential adverse effects.

Potential adverse effects include, but are not limited to, the following:

1. Loss of proper spinal curvature, correction, height, and/or reduction
2. Infection
3. Non-Union or delayed union
4. Foreign body reaction to the implants

5. Hemorrhaging
6. Loss of neurological function, dural tear, pain, and/or discomfort
7. Bone graft fracture, vertebral body fracture or discontinued growth of fused at, above and/or below the surgery level
8. Bending, loosening, fracture, disassembly, slippage and/or migration of the components
9. Revision surgery
10. Dysphagia
11. Bursitis
12. Bone loss and/or bone fracture due to stress shielding
13. Loss of bladder and/or bowel control
14. Injury to recurrent laryngeal nerve resulting in alteration of voice
15. Injury to esophagus and/or trachea
16. Death

#### **WARNINGS:**

The following are warnings for this device.

1. This device is not approved for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic, or lumbar spine.
2. Potential risks identified with the use of this device system, which may require additional surgery, include device component fracture, loss of fixation, non-union, fracture of the vertebrae, necrosis of the bone, neurological injury, and/or vascular or visceral injury.
3. The benefit of spinal fusion utilizing any cervical plating system has not been adequately established in patients with stable spines.
4. Patient selection and compliance will greatly affect the results. Patients suffering from obesity, malnutrition, and/or poor bone quality are poor candidates for spinal fusion. Patients who smoke or abuse alcohol are poor candidates for spinal fusion.
5. Patients who smoke should be advised of the consequences of the fact that an increased incidence of non-union has been reported with patients who smoke.
6. It is recommended that the locking rivets should only be engaged once, or disengaged once, if necessary.
7. The locking rivets should not be engaged until the surgeon has screwed and tightened all bone screws and is ready to close the soft tissues.
8. Failure to engage the locking rivets may increase the chances of screw back out from the plate if the screws become loose.
9. The implants and instruments are provided non-sterile and must be cleaned and sterilized before use. Device components should be sterilized using one of the noted validated sterilization cycle parameters.
10. A successful result is not always achieved in every surgical case due to many extenuating circumstances. This device is intended for temporary immobilization of the cervical spine in order to obtain a solid fusion mass using a bone graft.
11. Only surgeons trained and experienced in spinal decompression and bone grafting techniques should use the cervical plate. Preoperative and operating procedures, including knowledge of surgical techniques and proper selection and placement of the implants are essential considerations in the utilization of this device.
12. Do not reuse implants. Discard used, damaged, or otherwise suspect implants. AN IMPLANT SHOULD NEVER BE RE-USED. Any implant, once used, should be discarded. Even though it appears undamaged, it may have small defects and internal stress patterns that may lead to failure. Reuse can potentially compromise device performance and patient safety

#### **PREOPERATIVE:**

1. The surgeon should only consider utilizing the **Slimplicity® Solo Anterior Cervical Plate System** with those patients who meet the criteria in Indications.
2. The surgeon should avoid utilizing this device with those patients who have Contraindications.
3. The surgeon should make sure that all implants and instruments are unpacked, sterilized, and available prior to surgery.
4. The implants and instruments are provided non-sterile and must be cleaned and sterilized before use.
5. Implants and instruments should be inspected for surface flaws and scratches and should not be used in the presence thereof.
6. The surgeon should have a complete understanding of the surgical technique, design rationale, indications and contraindications.

#### **INTRAOPERATIVE:**

1. The surgical technique manual should be followed.

2. Damage to the nerves will cause loss of neurological functions. Extreme caution should be taken to avoid the spinal cord and nerve roots at all times.
3. When contouring the plate, the plate should not be reverse bent. Excessive and/or repeated bending of the plate should be avoided, as this can significantly weaken the plate.
4. Proper sizing and positioning of bone graft is essential to obtain successful spinal fusion. The bone graft must extend from the upper to the lower vertebrae to be used.
5. Notching and scratching of implants should be avoided.
6. Bone grafts should be used to insure stability.
7. The locking rivets and bone screws should be engaged and tightened firmly before closing soft tissue.

**POSTOPERATIVE:**

1. The **Slimplicity Solo Anterior Cervical Plate System** implants are designed and intended as temporary fixation implants. The implants should be removed after complete healing has occurred. Implants which are not removed after serving their intended purpose may bend, dislocate, or break and/or cause corrosion, localized tissue reaction, pain, infection, and/or bone loss due to stress shielding.
2. For best possible results, patients should be counseled to avoid lifting, twisting, physical activities, smoking, consuming alcohol, and any other activity that would compromise or delay the healing process.
3. The patient should be warned about the limitation of bending at the point of spinal fusion.
4. The removal of implants should be properly disposed of and are not to be reused under any circumstance.

**STERILIZATION**

The **Slimplicity® Solo Anterior Cervical Plate System** is supplied clean and non-sterile and must be sterilized prior to use. Remove all packaging before sterilization. Implants and instruments should be autoclave sterilized using one of the following validated cycle parameters.

**Note:** Flash sterilization is not recommended for the **Slimplicity Solo Anterior Cervical Plate System**.

Method	Cycle Type	Sterilization Temperature	Minimum Exposure Time
Steam	Pre-vacuum	270°F (132°C)	4 minutes

To assure maintenance of sterility we recommend:

- Utilization of a minimum drying time of 30 minutes in accordance with ANSI/AAMI ST79:2010, *Comprehensive guide to steam sterilization and sterility assurance in health facilities*.
- For USA: Use only FDA cleared sterilization wraps to enclose the sterilization tray.

**MAGNETIC RESONANCE ENVIRONMENT:**

The **Slimplicity Solo Anterior Cervical Plate System** has not been evaluated for safety and compatibility in the Magnetic Resonance environment. The **Slimplicity Solo Anterior Cervical Plate System** has not been tested for heating or migration in the Magnetic Resonance environment.

**STORAGE INSTRUCTIONS:**

All products should be stored in a cool, dry place.

**HOW SUPPLIED:**

The required components and specialized instruments are supplied non-sterile in a container suitable for steam sterilization or individually packaged as replacement product. All components and instruments may be purchased independently.

**CARE AND HANDLING:**

- All torque handles should be returned to the manufacturer for recalibration every six months.
- Please refer to ASTM standards such as F1744-96, "Standard Guide for Care and Handling of Stainless Steel Surgical Instruments" for additional information.
- Surgical instruments are subject to wear with normal usage. Instruments which have experienced extensive use or excessive force are susceptible to fracture. Surgical instruments should be used only for their intended purpose.
- **Precision Spine** recommends that all instruments be visually inspected for wear and disfigurement, as well as tested to ensure instruments are functioning properly prior to use. If instruments are discolored, have loose screws/pins, are out of alignment, are cracked or have other irregularities, **DO NOT USE**.

**CLEANING AND DECONTAMINATION:**

- These instructions are to be followed prior to initial use and reprocessing of the instruments.
- Reprocessing the instruments using the methods described herein, will not limit the useful life of the instruments. The useful life of the product is typically determined by wear and damage due to use.
- Transport trays should be considered reusable devices, and inspected for visible soils and must be cleaned.
- **WARNING:** The following Cleaning instructions have been validated. Failure to follow all steps may result in an improperly cleaned and sterilized instrument (Non-Sterile).
- **CAUTION:** In order to preserve optimal efficiency and safety of the instruments, the following instructions must be followed.
  - The use of metallic brushes, scrub pads or other articles that are likely to damage the instrument must be avoided.
  - Chemicals, such as chlorine or soda as well as organic or ammoniated acids or solvents (e.g. Acetone) that are likely to damage the instrument, must not be used.
  - Mercurial solutions are not recommended, as they corrode metal parts.
  - If applicable, disassemble instruments prior to Cleaning. Articulated instruments must be opened in order to allow the cleaning of all interstices.
  - Immediately after the surgical procedure, disallowing organic debris to dry on the instruments, remove as much debris as possible from each instrument using a water moistened gauze pad or wipe, exchanging the gauze pad or wipe as it becomes soiled. Do not allow organic debris to dry.
  - Prepare a neutral pH enzymatic cleaning solution per the manufacturer's instructions with warm tap water (35-40°C).
  - Immerse the instruments in the cleaning solution for a minimum of 10 minutes, activating any mechanisms 5X, so the enzymatic cleaner contacts all mated surfaces. Thoroughly scrub all instruments with a soft bristle cleaning brush while immersed in the enzymatic cleaning solutions. Be sure that thorough scrubbing also includes any lumens with an appropriately sized brush that contacts all surfaces. Change the soak solution after each utilization or if grossly soiled.
  - Rinse the instruments in warm tap water (35-40°C) for at least one minute.
  - Transfer the instruments into fresh enzymatic cleaning solution. Sonicate the instruments while immersed in the cleaning solution for a minimum of 15 minutes.
  - Thoroughly rinse all instruments and lumens with warm running water (35-40°C), for at least one minute each until flushing water runs clear. Use a hose or water jet to rinse any lumens, holes, or complex interfaces. Perform a second rinse with DI water, again using a hose or water jet to rinse any lumens, holes, or complex interfaces.
  - Dry with a sterile gauze, clean cloth and/or clean compressed air. Inspect instruments for cleanliness, function, and residual moisture. Any device that is not visually clean must be reprocessed.

#### **LUBRICATION:**

To protect instruments from staining and rusting during sterilization and storage, they should be lubricated with a water-soluble, preserved lubricant after each cleaning. Since effective ultrasonic cleaning removes all lubricant, relubrication is important. The lubricant should contain a chemical preservative to prevent bacterial growth in the lubricant bath. The bath solution should be made with demineralized water. A lubricant containing a rust inhibitor helps prevent electrolytic corrosion of points and edges. Immediately after cleaning, instrument should be immersed for 30 seconds and allowed to drain off, not wiped off. A lubricant film will remain through the sterilization to protect them during storage.

#### **SPECIAL NOTE FOR TORQUE LIMITING HANDLES:**

**(This note only applies to customers who purchase Torque Limiting Handles).**

The following are suggested guidelines for calibration cycles of Torque Limiting Handles. Note that these are general recommendations only and users are encouraged to determine specific calibration cycles for each product depending on their particular situation or usage. Return product after six months of use or, after 150 autoclave cycles or, after approximately 3000 actuations (Clicks) whichever comes first.

#### **MATERIAL SPECIFICATION:**

All components are made from medical grade stainless steel, titanium or titanium alloy described by such standards as ASTM F-138, ISO 5832-12, ASTM F-136 or ISO 5832-3. The products are supplied clean and "NON-STERILE".

:

#### **CLINICAL HISTORY**

These instructions for use are based upon current experience. The physician may wish to vary the procedure in accordance with clinical judgment.

**PRODUCT COMPLAINTS:**

Any complaint or dissatisfaction with product quality, performance, labeling, and/or safety should be reported to **Precision Spine®**. If any of the implants or instruments “malfunction” (i.e. do not meet any of their performance specifications or do not perform as intended) and/or are suspected to have caused or contributed to the death or serious injury of the patient, **Precision Spine** should be notified immediately by phone, fax or written correspondence. When filing a complaint, please provide the product description, product number, lot number, your name and address, and the nature of the complaint.

**ADDITIONAL INFORMATION:**

The surgical technique guide for the implantation of the **Slimplicity® Solo Anterior Cervical Plate System** is available upon request. If further information is required, please contact the manufacturer.



**Precision Spine, Inc.**  
 2050 Executive Drive  
 Pearl, MS 39208  
 USA  
 Phone: 1-601-420-4244  
 Toll Free: 1-888-241-4773  
 Fax: 1-601-420-5501



**Emergo Europe**  
 Prinsessegracht 20  
 2514 AP The Hague  
 The Netherlands

			<b>RX only</b>
SEE PACKAGE INSERT FOR LABELING LIMITATIONS	NOT STERILE	SINGLE USE ONLY	SALE BY PHYSICIAN PRESCRIPTION FOR USA ONLY
MANUFACTURED BY			